

VALIDATION

Tester Equipment Ustar Nucleic Acid Amplification analyzer	
Test Kit Easy NAT Novel Coronavirus (2019-nCoV) RNA	
COVID-19 Lot No. 2020042	Transport Medium
Expiration Date 10/20/2020	TM
TEST ID CV92-19	

(1) COMPARTOR NEGATIVE	
GenMark Instrument Number 1903100099	
Software Version 2.5.2.0	
Cabelt Huntington Hospital (CHH)	

UCR102 SN	Date	ID						Test Results										Re-test									
		Test Source	Sample	Specimen Types	EasyNAT Cartridge	Expiration of Cartridge	Clinical Symptoms	ORF1ab_Ct_N_CT	left leg Tt value	Right leg Tt value	ICL	ICR	Ustar Result	CONTROL	Biofire Torch Genmark eFlex (1)	Comments/Technician	Test Description	Date	ICL	ICR	ROX	Result	W.V. Lab	Comments	Technician		
210218070003A		Cov10	Q6262098	Nasopharyngeal	Cov102004210612885	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6262097	Nasopharyngeal	Cov102004210613915	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6270259	Nasopharyngeal	Cov102004210612805	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6270433	Nasopharyngeal	Cov102004210612492	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6270284	Nasopharyngeal	Cov102004210611811	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		JAS		Cov10										
210218070003A		Cov10	Q6270280	Nasopharyngeal	Cov102004210612700	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		JAS		Cov10										
210218070003A		Cov10	Q6270284	Nasopharyngeal	Cov102004210611911	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		JAS		Cov10										
210218070003A		Cov10	Q6270289	Nasopharyngeal	Cov102004210612700	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		SA		Cov10										
210218070003A		Cov10	Q6270282	Nasopharyngeal	Cov102004210610826	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		SA		Cov10										
210218070003A		Cov10	Q6262070	Nasopharyngeal	Cov102004210612642	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	q1518173	Nasopharyngeal	Cov102004210613042	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6260287	Nasopharyngeal	Cov102004210611443	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6261950	Nasopharyngeal	Cov102004210610216	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6262092	Nasopharyngeal	Cov102004210613038	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6310742	Nasopharyngeal	Cov102004210612856	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6301289	Nasopharyngeal	Cov102004210611450	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6310359	Nasopharyngeal	Cov102004210613029	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6311114	Nasopharyngeal	Cov102004210611842	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6310278	Nasopharyngeal	Cov102004210610125	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6301030	Nasopharyngeal	Cov102004210613039	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6300823	Nasopharyngeal	Cov102004210612331	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6310483	Nasopharyngeal	Cov102004210611192	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	R0010311	Nasopharyngeal	Cov102004210612374	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6311257	Nasopharyngeal	Cov102004210611853	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	R0010305	Nasopharyngeal	Cov102004210612371	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6261981	Nasopharyngeal	Cov102004210612354	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6290649	Nasopharyngeal	Cov102004210612421	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6291832	Nasopharyngeal	Cov102004210612366	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6291364	Nasopharyngeal	Cov102004210611802	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6291510	Nasopharyngeal	Cov102004210612354	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6291370	Nasopharyngeal	Cov102004210612343	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										
210218070003A		Cov10	Q6291808	Nasopharyngeal	Cov102004210612368	10/20/2020		N/A	N/A	N/A	N/A	Negative	Negative		TMS		Cov10										

(1) GenMark and Biofire have the EUA FDA authorization