

**VALIDATION**

Tester Equipment Ustar Nucleic acid Amplification analyzer  
 Test Kit Easy NAT Novel-Coronavirus (2019-nCoV) RNA

COVID-19 Lot N°	20200421	Transport Medium VTM
Expiration Date	10/20/2020	
TEST	COVID-19	

(1) <b>COMPARATOR NEGATIVE</b>
GenMarkDx instrument number 1903100395
Software Version : 2.5.2.0

UC0102 SN	Date	Test Source	Sample	ID		Clinical Symptoms	ORF1ab_Ct N_CT				Ustar Result	CONTROL		Technician	Test Description	Date	ORF1ab_Ct N_CT			Re-test				
							left leg Tt value	Right leg Tt value	ICL	ICR		Biofire Torch Genmark ePlex (1)	Comments				ICL	ICR	ROX	Result	W.V. Lab	Comments	Technician	
210218070003A	5/27/2020	CoV10	Q9262098	Nasopharyngeal	CoV102004210612885		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	5/27/2020	CoV10	Q9262027	Nasopharyngeal	CoV102004210610913		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	5/28/2020	CoV10	Q9270259	Nasopharyngeal	CoV102004210612805		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	5/28/2020	CoV10	Q9270433	Nasopharyngeal	CoV102004210612492		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	5/29/2020	CoV10	Q9270264	Nasopharyngeal	CoV102004210611811		N/A	N/A	N/A	N/A	Negative	Negative		JAS	CoV10									
210218070003A	5/29/2020	CoV10	Q9270280	Nasopharyngeal	CoV102004210612700		N/A	N/A	N/A	N/A	Negative	Negative		JAS	CoV10									
210218070003A	5/29/2020	CoV10	Q9270264	Nasopharyngeal	CoV102004210611811		N/A	N/A	N/A	N/A	Negative	Negative		JAS	CoV10									
210218070003A	5/29/2020	CoV10	Q9270280	Nasopharyngeal	CoV102004210612700		N/A	N/A	N/A	N/A	Negative	Negative		SA	CoV10									
210218070003A	5/29/2020	CoV10	Q9270262	Nasopharyngeal	CoV102004210610826		N/A	N/A	N/A	N/A	Negative	Negative		SA	CoV10									
210218070003A	5/29/2020	CoV10	Q9262070	Nasopharyngeal	CoV102004210612542		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	5/29/2020	CoV10	q1519173	Nasopharyngeal	CoV102004210613042		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	5/29/2020	CoV10	Q9290287	Nasopharyngeal	CoV102004210611443		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	6/1/2020	CoV10	Q9281950	Nasopharyngeal	CoV102004210610216		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	6/1/2020	CoV10	Q9290262	Nasopharyngeal	CoV102004210612638		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	6/1/2020	CoV10	Q9310742	Nasopharyngeal	CoV102004210612856		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	6/1/2020	CoV10	Q9301269	Nasopharyngeal	CoV102004210611450		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	6/1/2020	CoV10	Q9310350	Nasopharyngeal	CoV102004210613026		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	6/1/2020	CoV10	Q9311114	Nasopharyngeal	CoV102004210611842		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	6/1/2020	CoV10	Q9310278	Nasopharyngeal	CoV102004210610725		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	6/1/2020	CoV10	Q9301030	Nasopharyngeal	CoV102004210612939		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	6/2/2020	CoV10	Q9300823	Nasopharyngeal	CoV102004210612331		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	6/2/2020	CoV10	Q9310483	Nasopharyngeal	CoV102004210611182		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	6/2/2020	CoV10	R0010311	Nasopharyngeal	CoV102004210612374		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	6/2/2020	CoV10	Q9311257	Nasopharyngeal	CoV102004210611853		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	6/2/2020	CoV10	R0010305	Nasopharyngeal	CoV102004210612371		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	6/2/2020	CoV10	Q9291981	Nasopharyngeal	CoV102004210612354		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	6/2/2020	CoV10	Q9290649	Nasopharyngeal	CoV102004210612421		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	6/3/2020	CoV10	Q9291832	Nasopharyngeal	CoV102004210612366		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	6/3/2020	CoV10	Q9291384	Nasopharyngeal	CoV102004210611802		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	6/3/2020	CoV10	Q9291510	Nasopharyngeal	CoV102004210612354		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	6/3/2020	CoV10	Q9291370	Nasopharyngeal	CoV102004210612343		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									
210218070003A	6/3/2020	CoV10	Q9291808	Nasopharyngeal	CoV102004210612386		N/A	N/A	N/A	N/A	Negative	Negative		TMS	CoV10									

(1) GenMark and Biofire have the EUA FDA authorization